

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 7/12/17 Instrument Serial Number: 100374

Instrument Location: Hoonah Police Dept

Supervisor Name and Agency: William Mills Hoonah PD

Describe the Status Change or Issue with the Instrument:

taking out of service because touch screen is not
working

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 7/12/17 1600 hrs

For Use by SCDL

Additional Notes

Instrument was out of service with an expired tank in
February. When new tank was being installed touch screen
issues occurred.

BMB 8/8/17

Email completed form to _____

For questions contact the Breath Alcohol Program at 907-269-5740

Foster, Charles R (DPS)

From: Foster, Charles R (DPS)
Sent: Friday, June 02, 2017 3:11 PM
To: Roth-Frey, Nicolette A (DPS); Barnett, Brandi M (DPS); O'Bryant, Colleen S (DPS)
Cc: William Mills - Hoonah PD (bmills@cityofhoonah.org)
Subject: 100374 (Hoonah PD) Out of Service

Phone call from Bill Mills today:

The DMT touchscreen is not responding. He turned the DMT off and on again a few times and this did not fix the issue. I told him that we will send him a replacement in July (once the state passes a budget).

Charles Foster
Chemistry Supervisor
Scientific Director - Breath and Blood Alcohol Program
Alaska Scientific Crime Detection Laboratory
907-269-5760