

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 4/12/19 Instrument Serial Number: 100405  
100415 BMS  
3-22860 4/10/19

Instrument Location: DILLINGHAM PD 404 DST DILLINGHAM, AK 99576

Supervisor Name and Agency: TANER LOWERY DILLINGHAM POLICE DEPARTMENT

Describe the Status Change or Issue with the Instrument:

CRIME LAB EMAILED ME INFORMING ME THE OLD DATAMASTER HAS BEEN  
EXPERIENCING TANK INTERFERENCE + BLANK ERRORS

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 4/12/2019 1:00 pm

For Use by SCDL

Additional Notes

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740