

In Field Instrument Review
Date Range: 2022.05.31 – 2022.07.18
Review Completed on 2022.08.11 by Colleen O’Bryant

Status Messages and Action Taken

100344 (Juneau PD) – 2/35 Diagnostic Failed on Subject Test – Dry gas pressure was below 100psi. Tank change occurred on 2022.07.23. No action taken.

100346 (Kenai PD) – 1/15 Pump Error on Subject Tests, 1/15 Interference on Subject Tests – Both subject tests were followed by completed tests. No action taken.

100351 (Fairbanks PD) – 1/31 Blank Error on Subject Test – Instrument removed from service on 2022.07.04 for suspicion of 5 way valve issues (100351 2022.07.04 Change in Status).

100355 (Homer PD) – 1/26 Pump Error on Subject Tests, 1/26 Blank Error on Subject Tests - Both subject tests were followed by completed tests. No action taken.

100364 (Anchorage Airport PD) – 3/18 Interference on Subject Tests – All three on the same subject with elevated filter two readings and isopropanol and acetone detected in the blood alcohol sample. – No action taken.

100365 (Fairbanks PD) – 1/10 Pump Error on Subject Tests – Status message followed by a completed subject test – No action taken.

100367 (Seward PD) – 1/19 RFI on Subject Tests - Status message followed by a completed subject test – No action taken.

100373 (Anchorage PD) – 2/138 RFI on Subject Tests, 1/138 Pump Error on subject Test, 2/138 Blank Error on Subject Test – All status messages were followed by a completed test – No action taken.

100378 (Palmer PD) – 3/39 RFI on Subject Tests, 1/39 Pump Error on Subject Tests – Instrument removed from service on 2022.07.03 due to frequency of RFI status messages (100378 2022.07.03 Change In Status).

100380 (Ketchikan PD) – 1/12 Pump Error on Subject Tests - Status message followed by a completed subject test – No action taken.

100382 (Petersburg PD) – 1/2 Blank Error on Subject Tests – Test type “Other” – No action taken.

100385 (AST/Fairbanks) – 1/49 RFI on Subject Tests, 1/49 Pump Error on Subject Tests, 1/49 Interference on Subject Tests – RFI and Pump Error status messages were followed by completed tests. Interference status message had an elevated filter 2 reading. - No action taken

100387 (Palmer PD) – 2/11 RFI on Subject Tests – both status messages were on the same subject followed by a completed test. - No action taken.

100391 (AST/Willow) – 3/4 Standard Out of Range on Subject Tests – All status messages occurred on the same subject and followed by a completed test. Standard out of range was low, pressure in the tank between ~500-600psi. Tank change has not occurred. Recommend Monitoring.

100395 (Fairbanks Airport PD) – 1/2 Internal Standard Error on VOC – Internal standard average reading >3% - Recommend Replacement - Replacement mailed 2022.08.11.

100422 (JBER) – 9/37 RFI on Subject Tests, 3/37 Pump Errors on Subject Tests, 1/39 Tank Interference on Subject Tests – 2 Pump Errors and 1 Tank Interference occurred on the same subject test. JBER BTS was contacted regarding the status messages, specifically the RFIs. They communicated that they will instruct officers to turn off their radios in the vicinity of the instrument. (100422 2022.08.04 Change In Status) – Recommend Monitoring.

100423 (Ft. Wainwright MP) – 1/23 Pump Error on Subject Tests, 3/23 Interference on Subject Tests, 1/1 Tank Interference on VOC – Pump Error status message was followed by a completed test. All interference detected occurred on the same subject with an elevated filter two reading. Tank interference occurred during 2022.07.18 VOC. Successful VOC was completed on 2022.08.02. – No action taken.

100669 (AST/Soldotna) – 2/26 RFI on Subject Tests – No action taken.

100682 (Kodiak PD) – 1/8 Standard out of range on Subject Tests – Status message occurred during a subject test with an elevated alcohol level. Followed by a completed test. - No action taken.

100687 (Galena PD) – 1/4 Tank Interference on VOC, 1/4 Diagnostic Failed on VOC due to internal standard error – Both status messages were followed by two successful VOCs (2022.06.14, 2022.07.18) No subject tests performed during this time frame. Recommend Monitoring

100693 (Juneau PD) – 3/10 Blank Error on Subject Tests – Two blank errors occurred on the same subject with an elevated alcohol level. The subject was successfully tested on another instrument. Another blank error on an additional subject was followed by a completed test. Recommend Monitoring.

100694 (UAF PD) – 1/1 Tank Interference on VOC – Status message was followed by a successfully completed VOC. – No Action Taken.

100697 (Anchorage PD) – 1/138 Pump Error on Subject Tests, 1/138 Interference on Subject Tests, 2/138 Tank Interference on Subject Tests, 2/138 Blank Error on Subject Tests – Pump error, and blank error status messages were followed by completed tests. Tank interference occurred twice on the same subject showing slightly elevated filter 2 and filter 3 readings. Did not reoccur. - No action taken.

100699 (Valdez PD) – 2/12 RFI on Subject Tests, 1/12 Blank Error on Subject Tests, 2/12 Standard Out of Range on Subject Tests – Both RFI status messages were followed by completed tests. The Standard Out of Range and Blank Error status messages occurred on the same individual with an elevated alcohol concentration. These tests were followed by a completed test. – No action taken.

Follow Up from Previous In Field Instrument Review

100378 (Palmer PD): Due to the high prevalence of RFI status messages without a notable source (100378 2022.06.24 Change In Status), the instrument was replaced 2022.07.03 (100378 2022.07.03 Change In Status)

100683 (AST/Glennallen): Instrument removed from service on 2022.06.11 for lack of response from the touch screen (100683 2022.06.11 Change In Status)

Agencies not included in this summary: 100384 (Sand Point), 100424 (Eielson AFB), 100676 (Ft. Yukon PD)

Subject Test Data

Serial Number	Total Subject Tests	Date Range								Total Tests			Total Messages			Diagnostic Failc	Total Codes					
		RFI		Pump Error		Interference		Tank Interference		Blank Error		Standard Out of Range	Gas Flow Error	Ambient Fail	Internal Standard Error			Diagnostic Failc				
		5/31/2022	7/18/2022	5/31/2022	7/18/2022	5/31/2022	7/18/2022	5/31/2022	7/18/2022	5/31/2022	7/18/2022	5/31/2022	7/18/2022	5/31/2022	7/18/2022			5/31/2022	7/18/2022			
100344	35	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2	6%	2		
100346	15	0	0%	1	7%	1	7%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2
100351	31	0	0%	0	0%	0	0%	0	0%	1	3%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100355	26	0	0%	1	4%	0	0%	0	0%	1	4%	0	0%	0	0%	0	0%	0	0%	0	0%	2
100364	18	0	0%	0	0%	3	17%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	3
100365	10	0	0%	1	10%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100367	19	1	5%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100373	138	2	1%	1	1%	0	0%	0	0%	2	1%	0	0%	0	0%	0	0%	0	0%	0	0%	5
100378	39	3	8%	1	3%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	4
100380	12	0	0%	1	8%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100382	2	0	0%	0	0%	0	0%	0	0%	1	50%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100385	49	1	2%	1	2%	1	2%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	3
100387	11	2	18%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2
100391	4	0	0%	0	0%	0	0%	0	0%	0	0%	3	75%	0	0%	0	0%	0	0%	0	0%	3
100422	37	9	24%	3	8%	0	0%	1	3%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	13
100423	23	0	0%	1	4%	3	13%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	4
100669	26	2	8%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	2
100682	8	0	0%	0	0%	0	0%	0	0%	0	0%	1	13%	0	0%	0	0%	0	0%	0	0%	1
100693	10	0	0%	0	0%	0	0%	0	0%	3	30%	0	0%	0	0%	0	0%	0	0%	0	0%	3
100697	138	0	0%	1	1%	1	1%	2	1%	2	1%	0	0%	0	0%	0	0%	0	0%	0	0%	6
100699	12	2	17%	0	0%	0	0%	0	0%	1	8%	2	17%	0	0%	0	0%	0	0%	0	0%	5
Total	823	22	3%	12	1%	9	1%	3	0%	11	1%	6	1%	0	0%	0	0%	0	0%	2	0%	65

**The date range listed above represents the first and last dates of data generated from the DMHost report. The search range could be larger than the date range listed

Subject Test Data

Serial Number	Total Subject Tests	User Abort	Detector Overflow	Barometer Error	Filter Wheel Error	Date Range		Total Tests			Samp Chamb Low	Breath Tube High	Static Analysis Error	Total Codes
						5/31/2022	7/18/2022	Total Messages	823	0.0%				
100255	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100342	5	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100343	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100344	35	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100345	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100346	15	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100347	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100348	5	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100349	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100350	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100351	31	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100352	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100353	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100354	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100355	26	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100356	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100357	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100358	29	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100359	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100360	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100361	5	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100362	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100363	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100364	18	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100365	10	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100366	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100367	19	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100368	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100369	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100370	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100371	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100372	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100373	138	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100374	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100375	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100376	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100377	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100378	39	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100379	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100380	12	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100381	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100382	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100384	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100385	49	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100386	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100387	11	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100388	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100389	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100390	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100391	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100392	6	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100395	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100396	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100397	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100398	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100399	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100400	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100402	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100403	26	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100404	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100405	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100406	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100410	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100411	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100412	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100413	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100414	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100415	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100416	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100417	6	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100418	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100420	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100421	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100422	37	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100423	23	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100424	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100665	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100666	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100667	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100668	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100669	26	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100670	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100671	6	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100672	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100673	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100674	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100675	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100676	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100677	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100678	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100679	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100680	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100681	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100682	8	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100683	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100684	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100685	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100686	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100687	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100688	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100689	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100690	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100691	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100692	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100693	10	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100694	6	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100695	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100696	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100697	138	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100698	7	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100699	12	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100700	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
Total	823	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0

The date range listed above represents the first and last dates of data generated from the DMHost report. The search range could be larger than the date range listed.

VOC Data

Serial Number	Total VOC Tests	RFI		Pump Error		Interference		Tank Interference		Date Range		Total Tests				Total Messages		Ambient Fail	Internal Standard Error	Diagnostic Failed	Total Codes			
		0	0%	0	0%	0	0%	0	0%	0	0%	5/31/2022	7/18/2022	Blank Error	Standard Out of Range	Gas Flow Error	112					5	4.5%	
100395	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1	50%	0	0%	1
100423	1	0	0%	0	0%	0	0%	0	0%	1	100%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
100687	4	0	0%	0	0%	0	0%	0	0%	1	25%	0	0%	0	0%	0	0%	0	0%	0	0%	1	25%	2
100694	1	0	0%	0	0%	0	0%	0	0%	1	100%	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	1
Total	112	0	0%	0	0%	0	0%	0	0%	3	3%	0	0%	0	0%	0	0%	0	0%	1	1%	1	1%	5

The date range listed above represents the first and last dates of data generated from the DMHost report. The search range could be larger than the date range listed.

VOC Data

Serial Number	Total VOC Tests	User Abort	Detector Overflow	Barometer Error	Filter Wheel Error	Date Range		Total Tests			Samp Chamb Low	Breath Tube High	Static Analysis Error	Total Codes
						5/31/2022	7/18/2022	112	0	0.0%				
100255	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100342	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100343	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100344	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100345	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100346	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100347	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100348	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100349	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100350	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100351	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100352	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100353	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100354	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100355	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100356	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100357	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100358	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100359	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100360	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100361	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100362	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100363	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100364	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100365	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100366	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100367	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100368	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100369	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100370	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100371	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100372	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100373	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100374	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100375	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100376	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100377	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100378	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100379	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100380	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100381	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100382	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100384	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100385	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100386	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100387	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100388	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100389	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100390	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100391	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100392	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100395	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100396	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100397	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100398	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100399	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100400	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100402	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100403	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100404	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100405	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100406	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100410	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100411	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100412	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100413	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100414	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100415	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100416	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100417	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100418	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100420	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100421	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100422	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100423	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100424	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100665	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100666	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100667	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100668	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100669	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100670	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100671	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100672	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100673	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100674	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100675	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100676	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100677	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100678	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100679	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100680	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100681	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100682	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100683	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100684	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100685	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100686	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100687	4	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100688	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100689	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100690	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100691	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100692	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100693	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100694	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100695	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100696	1	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100697	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100698	3	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100699	2	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
100700	0	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0
Total	112	0	0%	0	0%	0	0%	0	0%	0	0%	0	0%	0

The date range listed above represents the first and last dates of data generated from the DMHost report. The search range could be larger than the date range listed.