

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMCIF 2015 R1  
Status: Active

Instrument S/N 100397

Received From King Cove

Date 8/9/16

Analyst Brandi Barnett

1. Reason for instrument return:

Instrument would not upload. Unsure if it was due on instrument.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From NPAS

7. NPAS Repair Paperwork (attached)