



State of Alaska  
 Department of Public Safety  
**Scientific Crime Detection Laboratory**

4805 Dr. Martin Luther King Jr. Ave.  
 Anchorage, Alaska 99507  
 David Kanaris, Chief, Forensic Laboratories

Breath Alcohol Calibration Report  
 Calibration Item: Datamaster DMT  
 Serial Number: 100345  
 Date Received: Not Recorded  
 Condition of item: Not Recorded  
 Method Used: DataMaster DMT Calibration/Current BAPM  
 Calibration Range: 0.020 g/210L - 0.300 g/210L

Report Date: 10/27/2022  
 K690AGCTNJXZTP

**No As-Found Data Collected**

**Start Date 10/12/2022 - As Left - End Date 10/13/2022**

As Left Conditions - Barometric Pressure (inHg): 29.51 Temperature (C): 22.3 Relative Humidity (%): 20

**As-Left Data**

Nominal Values (g/210L)	Target Value at 29.51 in Hg	Lot Number	Results (g/210L)					Mean (g/210L)	Uncertainty (g/210L)
0.020	0.020	AG100401	0.020	0.020	0.021	0.020	0.020	0.020	+/- 0.002
0.040	0.039	AG111710	0.039	0.039	0.039	0.039	0.039	0.039	+/- 0.003
0.080	0.079	AG113402	0.078	0.077	0.077	0.078	0.078	0.078	+/- 0.003
0.100	0.099	AG211003	0.097	0.097	0.097	0.097	0.097	0.097	+/- 0.003
0.150	0.148	AG134205	0.147	0.145	0.145	0.146	0.146	0.146	+/- 0.004
0.300	0.295	AG217903	0.293	0.292	0.292	0.292	0.292	0.292	+/- 0.008

No As-Found Evaluation was performed. The Evaluation performed prior to this Calibration was completed before Breath Alcohol Procedure Manual version 5.0.

Calibration is performed using dry gas standards traceable to NIST. Standards are sourced from an external provider accredited to ISO/IEC 17025 and ISO 17034 standards. Internal instrument barometer readings are traceable to NIST. External barometers are calibrated by an external calibration provider accredited to the ISO/IEC 17025 standard.

The As-Left results listed above include the estimated measurement uncertainty with a coverage factor k=2 and an expanded coverage probability of 95%. The estimated uncertainty applies to the calibration process and not to subject breath samples.

**This report applies only to the item being calibrated and shall not be reproduced, except in full, without written approval from the laboratory.**

Analyst: Colleen O'Bryant  
[colleen.obryant@alaska.gov](mailto:colleen.obryant@alaska.gov)

Technically Reviewed by: Brandi Barnett on 10/27/2022

# Alaska Scientific Crime Detection Laboratory

## DataMaster Calibration Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100345 Analyst: C. O. Bryant Start Date: 10/12/22

- Diagnostic Test (attached)
- Adjust voltages, if needed
  - Lamp (2.60 V Maximum)
  - Bias
  - Cooler (2.00 V Maximum)
  - Chopper Frequency
  - Detector Voltage (-0.100V to 0.100V)
- Internal Standard value  $\leq$  3.00%
- Set RFI Sensitivity
- Adjust internal barometer reading to match NIST-traceable barometer
- Adjust volume setting, if needed (1.50 L +/- 0.10 L)
- Print Technician Screen (attached)
- Print Calibration Factors (attached)
- Minimum Volume Test 1.5 L (1.3L to 1.7L)

As Left Conditions:

Barometric Pressure (in Hg): ~~29.92~~ Temperature (C): 22.3 Relative Humidity (%) 20

Linearity Test (attached) 29.51  
COB 2022.10.17

0.020g/210L Mnf:	<u>Aurgas</u>	Lot:	<u>AG100401</u>	Exp:	<u>1/4/23</u>
0.040g/210L Mnf:	<u>Aurgas</u>	Lot:	<u>AG111710</u>	Exp:	<u>4/27/23</u>
0.080g/210L Mnf:	<u>Aurgas</u>	Lot:	<u>AG113402</u>	Exp:	<u>5/14/23</u>
0.100g/210L Mnf:	<u>Aurgas</u>	Lot:	<u>AG211003</u>	Exp:	<u>4/20/24</u>
0.150g/210L Mnf:	<u>Aurgas</u>	Lot:	<u>AG134205</u>	Exp:	<u>12/08/23</u>
0.300g/210L Mnf:	<u>Aurgas</u>	Lot:	<del>AG130502</del> <u>AG217903</u>	Exp:	<del>4/1/23</del> <u>6/28/24</u>

Functional Tests (attached)

- Incomplete
- Maximum Attempts Exceeded
- Interference (Volatiles)
- Invalid
- Radio Frequency Interference (RFI)

Acetone Mnf: Lifeloc Lot: J1DP41158 Level: 0.100 Exp: 4/09/23

Simulator S/N: SD200 Verification Performed?

Isopropanol Mnf: Lifeloc Lot: J1KP41172 Level: 0.100 Exp: 11/9/23

Simulator S/N: DR4176 Verification Performed?  11/9/23

Neg. Control Mnf: Lifeloc Lot: J1KP41172 Level: 0.100 Exp: 4/2/23

Simulator S/N: DR4162 Verification Performed?

Calibrated External Thermometer:

Thermometer S/N: 210958384 Calibration Due Date: 11/9/23

- Verify stylus is calibrated
- Reset Options
- Clear Memory
- Verification of Calibration (VOC)
- Non-Drinking Subject Test
- Upload
- Remove Tank Prior to Shutdown
- Calibration End Date: 10/13/22

DIAGNOSTIC RECORD

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 13:32:48

VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED  
(44.0°C - 52.0°C)  
Breath Tube = 45.7°C PASSED  
(38.0°C - 50.0°C)

SETTINGS

Lamp Voltage = 1.70 V PASSED  
(1.02 V - 2.38 V)  
Cooler Voltage = 1.64 V PASSED  
(0.98 V - 2.30 V)  
Bias Voltage = 80 V PASSED  
(48 V - 112 V)  
Chopper Freq = 532 Hz PASSED  
(475 Hz - 575 Hz)  
Barometer = 29.5 in

PUMP INFO

Flow Rate = 4.311 L/M PASSED  
(3.500 L/M - 6.500 L/M)

DETECTOR INFO

PUMP ON PASSED  
(0.002 V <= 0.010 V)  
PUMP OFF PASSED  
(0.002 V <= 0.010 V)

FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD

Xq = 0.093 1.02% PASSED  
(0.00% - 4.00%)

Technician Screen

-----  
DATAMASTER dmt: 100345  
-----

Date: 10/12/2022  
Time: 1:42 PM

Temperatures (°C)  
Sample Cell: 48.8°C  
Breath Tube: 45.5°C

Current Barometer 29.5 in  
Volume (Ltr) 1.49

Settings  
Lamp: 1.72 V  
Bias: 80 V  
Cooler: 1.66 V  
Chopper: 531 Hz

Voltages (V)  
Flow: 1.06 V  
Detector: 0.056 V

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100345  
-----

Date: 09/01/2022  
Time: 09:10:46

OPERATOR NAME:  
TATIANA MALNOSKE  
CERTIFICATE #: 10437

Ca = 0.1000  
CAL = 0.970856    0.800 <= CAL < 1.200  
b1 = 0.0014    0.0000 <= b1 < 0.0040  
b2 = 0.0059    0.0010 <= b2 < 0.0100  
b3 = 0.0008    0.0000 <= b3 < 0.0040  
Xq = 0.0945    0.0500 <= Xq < 0.2000  
a21 = 1.146871    1.050 <= a21 < 1.300  
a31 = 0.502084    0.300 <= a31 < 0.800

SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 13:43:43


SUBJECT NAME:  
TEST TEST  
LICENSE #: VOLUME TEST

OPERATOR NAME:  
COLLEEN S OBRYANT  
OPERATOR'S #: 6237

DEPT/AGENCY:  
AAA1  
CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION  
NOMINAL: 0.080  
TARGET AT 29.51: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK PASSED 13:44  
BLANK TEST 0.000 13:44  
INTERNAL STANDARD VERIFIED 13:45  
EXTERNAL STANDARD 0.078 13:45  
BLANK TEST 0.000 13:46  
SUBJECT SAMPLE (Vol=1.52L) 0.000 13:47  
BLANK TEST 0.000 13:48  
EXTERNAL STANDARD 0.078 13:48  
BLANK TEST 0.000 13:49



LINEARITY CHECK


Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 13:53:30

SUPERVISOR NAME:  
COLLEEN S O'BRYAN  
OPERATOR'S #: 6237

TARGET	AVERAGE	STD. DEV.
0.020	0.0202	0.0004
0.039	0.0390	0.0000
0.079	0.0776	0.0005
0.099	0.0970	0.0000
0.148	0.1458	0.0008
0.295	0.2922	0.0004

SLOPE = 0.989426  
R<sup>2</sup> = 0.999961



CVB

LINEARITY CHECK: Control Test 1

Alaska Department of Public Safety  
DATAMASTER dmt: 100345






Date: 10/12/2022  
Time: 13:53:32

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020  
TARGET AT 29.51: 0.020  
LOT #: AG100401  
EXPIRATION: 01/04/2023  
TANK PRESSURE: 314 psi

BLANK TEST	0.000	13:54
INTERNAL STANDARD	VERIFIED	13:54
EXTERNAL STANDARD	0.020	13:55
BLANK TEST	0.000	13:56
EXTERNAL STANDARD	0.020	13:56
BLANK TEST	0.000	13:57
EXTERNAL STANDARD	0.021	13:57
BLANK TEST	0.000	13:58
EXTERNAL STANDARD	0.020	13:58
BLANK TEST	0.000	13:59
EXTERNAL STANDARD	0.020	13:59
BLANK TEST	0.000	14:00

Average = 0.0202  
Std Dev = 0.0004





LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety  
 DATAMASTER dmt: 100345






Date: 10/12/2022  
 Time: 14:02:45

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.040  
 TARGET AT 29.51: 0.039  
 LOT #: AG111710  
 EXPIRATION: 04/27/2023  
 TANK PRESSURE: 504 psi

BLANK TEST	0.000	14:04
INTERNAL STANDARD	VERIFIED	14:04
EXTERNAL STANDARD	0.039	14:04
BLANK TEST	0.000	14:05
EXTERNAL STANDARD	0.039	14:05
BLANK TEST	0.000	14:06
EXTERNAL STANDARD	0.039	14:07
BLANK TEST	0.000	14:07
EXTERNAL STANDARD	0.039	14:08
BLANK TEST	0.000	14:09
EXTERNAL STANDARD	0.039	14:09
BLANK TEST	0.000	14:10

Average = 0.0390  
 Std Dev = 0.0000



LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety  
 DATAMASTER dmt: 100345






Date: 10/12/2022  
 Time: 14:11:00

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
 TARGET AT 29.50: 0.079  
 LOT #: AG113402  
 EXPIRATION: 05/14/2023  
 TANK PRESSURE: 941 psi

BLANK TEST	0.000	14:12
INTERNAL STANDARD	VERIFIED	14:12
EXTERNAL STANDARD	0.078	14:12
BLANK TEST	0.000	14:13
EXTERNAL STANDARD	0.077	14:13
BLANK TEST	0.000	14:14
EXTERNAL STANDARD	0.077	14:14
BLANK TEST	0.000	14:15
EXTERNAL STANDARD	0.078	14:15
BLANK TEST	0.000	14:16
EXTERNAL STANDARD	0.078	14:16
BLANK TEST	0.000	14:17

Average = 0.0776  
 Std Dev = 0.0005



LINEARITY CHECK: Control Test 4

Alaska Department of Public Safety  
DATAMASTER dmt: 100345






Date: 10/12/2022  
Time: 14:19:10

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.100  
TARGET AT 29.49: 0.099  
LOT #: AG211003  
EXPIRATION: 04/20/2024  
TANK PRESSURE: 1112 psi

BLANK TEST	0.000	14:21
INTERNAL STANDARD	VERIFIED	14:21
EXTERNAL STANDARD	0.097	14:21
BLANK TEST	0.000	14:22
EXTERNAL STANDARD	0.097	14:22
BLANK TEST	0.000	14:23
EXTERNAL STANDARD	0.097	14:23
BLANK TEST	0.000	14:24
EXTERNAL STANDARD	0.097	14:24
BLANK TEST	0.000	14:25
EXTERNAL STANDARD	0.097	14:25
BLANK TEST	0.000	14:26

Average = 0.0970  
Std Dev = 0.0000



LINEARITY CHECK: Control Test 5






Alaska Department of Public Safety  
 DATAMASTER dmt: 100345

Date: 10/12/2022  
 Time: 14:27:42

EXTERNAL STANDARD INFORMATION  
 NOMINAL: 0.150  
 TARGET AT 29.48: 0.148  
 LOT #: AG134205  
 EXPIRATION: 12/08/2023  
 TANK PRESSURE: 1041 psi

BLANK TEST	0.000	14:28
INTERNAL STANDARD	VERIFIED	14:29
EXTERNAL STANDARD	0.147	14:29
BLANK TEST	0.000	14:30
EXTERNAL STANDARD	0.145	14:30
BLANK TEST	0.000	14:31
EXTERNAL STANDARD	0.145	14:31
BLANK TEST	0.000	14:32
EXTERNAL STANDARD	0.146	14:32
BLANK TEST	0.000	14:33
EXTERNAL STANDARD	0.146	14:33
BLANK TEST	0.000	14:34

Average = 0.1458  
 Std Dev = 0.0008



LINEARITY CHECK: Control Test 6

Alaska Department of Public Safety  
 DATAMASTER dmt: 100345






Date: 10/12/2022  
 Time: 14:53:18

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300  
 TARGET AT 29.44: 0.295  
 LOT #: AG217903  
 EXPIRATION: 06/28/2024  
 TANK PRESSURE: 555 psi

BLANK TEST	0.000	14:54
INTERNAL STANDARD	VERIFIED	14:54
EXTERNAL STANDARD	0.293	14:54
BLANK TEST	0.000	14:55
EXTERNAL STANDARD	0.292	14:56
BLANK TEST	0.000	14:57
EXTERNAL STANDARD	0.292	14:57
BLANK TEST	0.000	14:58
EXTERNAL STANDARD	0.292	14:58
BLANK TEST	0.000	14:59
EXTERNAL STANDARD	0.292	14:59
BLANK TEST	0.000	15:00

Average = 0.2922  
 Std Dev = 0.0004



SUBJECT TEST

Alaska Department of Public Safety

DATAMASTER dmt: 100345

Date: 10/12/2022

Time: 16:24:52

SUBJECT NAME:

TEST TEST

LICENSE #: INCOMPLETE

OPERATOR NAME:

COLLEEN S O'BRYAN

OPERATOR'S #: 6237

DEPT/AGENCY:

AAA1

CASE/REPORT: 22-1234

TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.37: 0.079

LOT #: AG113402

EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK	PASSED	16:27
BLANK TEST	0.000	16:27
INTERNAL STANDARD	VERIFIED	16:27
EXTERNAL STANDARD	0.077	16:27
BLANK TEST	0.000	16:28
SUBJECT SAMPLE	INCOMPLETE	16:30

— Alcohol (g/210L)  
 - - - - Flow Rate (L/M)  
 Min. Flow Rate (L/M)



SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 16:33:51

SUBJECT NAME:  
TEST TEST  
LICENSE #: INVALID


OPERATOR NAME:  
COLLEEN S O'BRYAN  
OPERATOR'S #: 6237

DEPT/AGENCY:  
AAA1  
CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION  
NOMINAL: 0.080  
TARGET AT 29.36: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK	PASSED	16:34
BLANK TEST	0.000	16:35
INTERNAL STANDARD	VERIFIED	16:35
EXTERNAL STANDARD	0.077	16:35
BLANK TEST	0.000	16:36
SUBJECT SAMPLE	INVALID	16:36

*Handwritten signature*



SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/13/2022  
Time: 06:56:32

SUBJECT NAME:  
TEST TEST  
LICENSE #: MAX ATTEMPTS


OPERATOR NAME:  
COLLEEN S OBRYANT  
OPERATOR'S #: 6237

DEPT/AGENCY:  
AAA1  
CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION  
NOMINAL: 0.080  
TARGET AT 29.41: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK PASSED 06:57  
BLANK TEST 0.000 06:57  
INTERNAL STANDARD VERIFIED 06:57  
EXTERNAL STANDARD 0.077 06:57  
BLANK TEST 0.000 06:58  
SUBJECT SAMPLE XXX 06:59

Maximum Attempts Exceeded





SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 16:39:22

SUBJECT NAME:  
TEST TEST  
LICENSE #: RFI

OPERATOR NAME:  
COLLEEN S O'BRYANT  
OPERATOR'S #: 6237


DEPT/AGENCY:  
AAA1  
CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION  
NOMINAL: 0.080  
TARGET AT 29.36: 0.078  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK PASSED 16:41  
BLANK TEST 0.000 16:41  
INTERNAL STANDARD VERIFIED 16:41  
EXTERNAL STANDARD 0.077 16:41  
BLANK TEST 0.000 16:42

RFI DETECTED

— Alcohol (g/210L) -  
- - - - Flow Rate (L/M)  
- - - - Min. Flow Rate (L/M)



SUBJECT TEST

Alaska Department of Public Safety

DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 15:37:09

SUBJECT NAME:  
TEST TEST  
LICENSE #: INT-ACETONE

OPERATOR NAME:  
COLLEEN S OBYRANT  
OPERATOR'S #: 6237

DEPT/AGENCY:  
AAA1  
CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.42: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023


DIAGNOSTIC CHECK	PASSED	15:38
BLANK TEST	0.000	15:38
INTERNAL STANDARD	VERIFIED	15:38
Xg = 0.094 (0.74%)		
EXTERNAL STANDARD	0.077	15:39
X[1] = 0.0774 (-0.0014) (0.0020)		
BLANK TEST	0.000	15:40
SUBJECT SAMPLE	xxx	15:40
X[1] = 0.1079 (-0.0163) (0.0023)		

Interference Detected

Filter 2 Avg = -0.0010

Filter 3 Avg = 0.0025

COB 2022.10.17



SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 15:45:47

SUBJECT NAME:  
TEST TEST  
LICENSE #: INT-IPA



OPERATOR NAME:  
COLLEEN S O'BRYAN  
OPERATOR'S #: 6237

DEPT/AGENCY:  
AAA1  
CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION  
NOMINAL: 0.080  
TARGET AT 29.40: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK	PASSED	15:46
BLANK TEST	0.000	15:46
INTERNAL STANDARD	VERIFIED	15:47
$X_q = 0.094$ (0.65%)	0.077	15:47
EXTERNAL STANDARD	$X[1] = 0.0770$ (-0.0013)	(0.0024)
BLANK TEST	0.000	15:48
SUBJECT SAMPLE	XXX	15:48
	$X[1] = 0.1157$ (-0.0260)	(0.0008)

Interference Detected



SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/12/2022  
Time: 15:28:39

SUBJECT NAME:  
TEST TEST  
LICENSE #: INT-NEGATIVE

OPERATOR NAME:  
COLLEEN S O'BRYAN  
OPERATOR'S #: 6237

DEPT/AGENCY:  
AAA1


CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.43: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK	PASSED	15:29
BLANK TEST	0.000	15:30
INTERNAL STANDARD	VERIFIED	15:30
$X_q = 0.094$ (0.37%)		
EXTERNAL STANDARD	0.078	15:30
$X[1] = 0.0775$ (-0.0007)	(0.0028)	
BLANK TEST	0.000	15:31
SUBJECT SAMPLE (Vol=2.05L)	0.092	15:32
$X[1] = 0.0924$ (-0.0010)	(0.0009)	
BLANK TEST	0.000	15:33
EXTERNAL STANDARD	0.077	15:33
$X[1] = 0.0773$ (-0.0007)	(0.0027)	
BLANK TEST	0.000	15:34

— Alcchol(g/210L) . . . .  
 - - - - Flow Rate(L/M)  
 . . . . Min. Flow Rate(L/M)



SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100345

Date: 10/13/2022  
Time: 07:11:08

SUBJECT NAME:  
TEST TEST  
LICENSE #: NON DRINKING

OPERATOR NAME:  
COLLEEN S OBRYANT  
OPERATOR'S #: 6237

DEPT/AGENCY:  
AAA1  
CASE/REPORT: 22-1234  
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION  
NOMINAL: 0.080  
TARGET AT 29.42: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK	PASSED	07:12
BLANK TEST	0.000	07:12
INTERNAL STANDARD	VERIFIED	07:12
EXTERNAL STANDARD	0.077	07:12
BLANK TEST	0.000	07:13
SUBJECT SAMPLE (Vol=2.17L)	0.000	07:14
BLANK TEST	0.000	07:15
EXTERNAL STANDARD	0.078	07:15
BLANK TEST	0.000	07:16

