

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100685

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/15/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.82: 0.080
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 635 psi

| | | |
|-------------------|--------------|-------|
| BLANK TEST | 0.000 | 13:01 |
| INTERNAL STANDARD | VERIFIED | 13:01 |
| EXTERNAL STANDARD | INTERFERENCE | 13:01 |

Blank Error

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 48.9°C | PASSED |
| Breath Tube = 43.3°C | PASSED |

PUMP INFO
Flow Rate = 5.648 L/M PASSED

DETECTOR INFO
PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO
Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File only. Instrument taken out of service on 02/15/2019 for multiple status messages. JW 03/15/2019