

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100414

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/08/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.31: 0.081
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 1001 psi

BLANK TEST	0.000	15:23
INTERNAL STANDARD	VERIFIED	15:24
EXTERNAL STANDARD	0.080	15:24
BLANK TEST	0.000	15:25
EXTERNAL STANDARD	0.080	15:25
BLANK TEST	0.000	15:26
EXTERNAL STANDARD	0.081	15:26
BLANK TEST	0.000	15:27
EXTERNAL STANDARD	INTERFERENCE	15:27

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 43.5°C	PASSED

PUMP INFO

Flow Rate = 4.917 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------