

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMCIF 2015 R1  
Status: Active

Instrument S/N 100365

Received From NPAS

Date 10/23/15

Analyst Brandi Barnett

1. Reason for instrument return:

Return from NPAS

2. Change instrument and case location in JusticeTrax Breath Alcohol Database
3. Ensure change in instrument status form is in JusticeTrax  N/A
4. Upload
5. Update agency information and polling status in DM Host Calling Program
6. Check for VOCs  N/A

From NPAS

7. NPAS Repair Paperwork (attached)



Work Ord No 49971  
Date 9/23/15

National Patent Analytical Systems, Inc.  
Service Action Report

Serial No 100365

Customer Alaska Dept of Public Phone 907-269-5582 Contact Bolz

War Exp 1/13/13 Mx Exp Ship Damage?  Y  N Cust Pick Up??  Y  N

Initial Comments: note any shipping damage or estimates required, along with accessories

Describe problem seen and/or attach documents received see attached document

Keyword

Finding Error Code Enter Hours 2.0

Repair Comments and work Performed updated software,evaluated and ran several accuracy tests with no problems

Solution Lot No

Voltage Set Values In				Voltage Set Values Out			
Lamp		Brth Vol		Lamp		Brth Vol	
Meter		Filter Dif		Meter		Filter Dif	
Stability				Stability			

Part No	Qty Ord	Pick	Back Ordered	Total ship	Description	Price Ea
9DMT	1.0				DATAMASTER DMT SERVICE RETURN 100365	
NOTE	1.0				NANUK CASE	
r500-1	2.0				Service Charge for Repair warranty	0.000

Date Completed 10/12/15

Technician M. BARBELY

Pick up Signature