

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: AUGUST 31, 2018 Instrument Serial Number: 100395

Instrument Location: 359 RAVIN DRIVE, HOONAH ALASKA

Supervisor Name and Agency: LT. WILLIAM MILLS, HOONAH POLICE DEPARTMENT

Describe the Status Change or Issue with the Instrument:

I REATTACHED THE CHECK VALVETO THE SIDE PANEL AND ALSO RETTACHED ON OF THE HOSES THAT HAD COME C
I DID A VCO AND PUT THE INSTRUMENT BACK IN SERVICES

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: AUGUST 31, 2018 1430 HRS

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740