

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100421

Date: 05/02/2019
Time: 07:30:16

OPERATOR NAME:
INTOX I I
CERTIFICATE #: I

Ca = 0.2000
CAL = 0.985310 0.800 <= CAL < 1.200
b1 = 0.0013 0.0000 <= b1 < 0.0040
b2 = 0.0035 0.0010 <= b2 < 0.0100
b3 = 0.0008 0.0000 <= b3 < 0.0040
Xq = 0.0992 0.0500 <= Xq < 0.2000
a21 = 1.125463 1.050 <= a21 < 1.300
a31 = 0.461934 0.300 <= a31 < 0.800