

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100367

Received From Intoximeters

Date 09/30/2019

Analyst Derck Walton

1. Reason for instrument return:

factory refurbish - filter wheel errors

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax  NA

4. Upload  NA

5. Update agency information and polling status in DM Host Calling Program  NA

6. Check for VOCs  NA

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)

**SRO Number:** SRO-036627

**Received Date:** 09/16/2019

**Completion Date:** 09/24/2019

**SRO Type:** REPAIR

**SRO Description:** DMT Repair/Refurb

**Customer ID:** C000003651

**Customer PO:** 09162019-Barnett

**Contact:**

**Email:**

**Ship To:**

**Units on SRO**

Alaska Department of Public Safety  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

1 LEGDMT LEGACY DMT

<b>Service Line: 1</b>		<b>**All Instruments are Calibrated to Factory Specifications**</b>
<b>Unit Type:</b>	(LEGDMT) LEGACY DMT	
<b>Serial Number:</b>	50100367	
<b>Reason:</b>	DMT Status Messages	
	FltWl er	
<b>Reason Notes:</b>	Customer Reported Issue: Filter wheel errors. Chopper wheel motor not functioning. Perform refurbishment and regulator replacement	
<b>Resolution:</b>	Replaced	
	RPL comp	
<b>Resolution Notes:</b>	90-8000-00 Prev Maint Contract p/n Replaced Dry Gas regulator Replaced Control Board Rebuilt filter assembly and added epoxy to motor wires Optimized detector Adjusted tech screen settings and RFI sensitivity Added o-ring to end of breath tube Adjusted Time and Date  001217-00 regulator 59-0006-00 Chopper motor 001199-00 3.3volt regulator 59-0064-00 Control board  Calibration solution (0.200) Lot#19040 Bottle#78 Dry gas: AG916303, Tank 011, Exp. 12 Jun 2021	

<b>Parts:</b>	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	59-0006-00	DMT HRDW ASSY MOTOR 12X20.6MM 41310
	1	001217-00	DMT HRDW GAS MOUNT ASSY W/VALVE
	1	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN
	1	59-0064-00	DMT ELEC PCB ASSY CONTROLLER 2010 41631
	1	90-8000-00	CONTRACT PREV MAINT DMT

# Order Packing Slip Report

**Intoximeters**  
Experience • Service • Integrity

9/25/2019 2:01:55 PM

**Packing Slip:** 186863

**Pack Date:** 9/25/2019

**SRO:** SRO-036627



**From:** Intoximeters, Inc  
2081 Craig Road  
St Louis MO 63146  
USA

**Customer:** C000003651

**Ship To:** 1

Alaska Department of Public Safety  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

Alaska Scientific Crime Detection  
Laboratory  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

**Customer PO:** 09162019-Barnett

**Ship Via:** FDX

FED EX

**Weight:** 38.00

**Delivery Terms:** 2D

**FOB:** ORIGIN

**# Packages:** 1

**Customer Ship Account:**

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100367
1	59-0006-00	EA		1.00	0.00
	DMT HRDW ASSY MOTOR 12X20.6MM 41310				
2	001217-00	EA		1.00	0.00
	DMT HRDW GAS MOUNT ASSY W/VALVE				
3	001199-00	EA		1.00	0.00
	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN				
4	59-0064-00	EA		1.00	0.00
	DMT ELEC PCB ASSY CONTROLLER 2010 41631				
5	90-8000-00	EA		1.00	0.00
	CONTRACT PREV MAINT DMT				