

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100685

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/15/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.82: 0.080
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 631 psi

BLANK TEST	0.000	13:05
INTERNAL STANDARD	VERIFIED	13:05
EXTERNAL STANDARD	INTERFERENCE	13:05

Interference Detected

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 41.7°C	PASSED

PUMP INFO

Flow Rate = 5.579 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File only, Instrument taken out of
service on 02/15/2019 for multiple
status messages. JMW 03/15/2019