

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 9/10/15 Instrument Serial Number: 100696

Instrument Location: Hooper Bay

Supervisor Name and Agency: Brandi Barnett SCDL

Describe the Status Change or Issue with the Instrument:

Instrument not passing 9/5 VOC. Interference detected on ~~internal standard~~ external standard & internal standard ^{EMB 9/10} error. Contacted James Heelsler and had him remove instrument from service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 9/10/15 9:30 AM

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740