

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100405

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/18/2013

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.36: 0.078
LOT #: 02012080A1
EXPIRATION: 03/01/2014

BLANK TEST	0.000	08:13
INTERNAL STANDARD	VERIFIED	08:13
EXTERNAL STANDARD	0.079	08:13
BLANK TEST	0.000	08:14
EXTERNAL STANDARD	0.078	08:14
BLANK TEST	0.000	08:15
EXTERNAL STANDARD	0.079	08:15
BLANK TEST	0.000	08:16
EXTERNAL STANDARD	0.078	08:17
BLANK TEST	0.000	08:17
EXTERNAL STANDARD	0.078	08:18
BLANK TEST	0.000	08:19

Average = 0.0784
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 2.00
PIC: 3.02
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED
Breath Tube = 46.5°C PASSED

PUMP INFO

Flow Rate = 4.749 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO


Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

4/5/2013

Subscribed and sworn before me this 5th day of April, 20 13


Nikki Roth, Notary Public
My Commission Expires With Office



CB 4/1/13