

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100374

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date:09/14/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.92: 0.080  
LOT #: AG634001  
EXPIRATION: 12/05/2018  
TANK PRESSURE: 103 psi

BLANK TEST	0.000	22:33
INTERNAL STANDARD	VERIFIED	22:33
EXTERNAL STANDARD	0.077	22:33
BLANK TEST	0.000	22:34

Tank Pressure Too Low

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 43.6°C	PASSED

### PUMP INFO

Flow Rate = 4.226 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED