



THE STATE
of **ALASKA**
GOVERNOR BILL WALKER

Department of Public Safety

SCIENTIFIC CRIME DETECTION LABORATORY

4805 Dr. Martin Luther King Jr. Avenue
Anchorage, Alaska 99507-1225
Main: 907.269.5740
Fax: 907.868.7946

Anchorage Jail Datamaster DMTs 100351 and 100374

June 3, 2016

The State of Alaska, Scientific Crime Detection Breath Alcohol program is aware of the following information regarding DataMaster DMTs with serial numbers 100351 and 100374, instruments that had been located at the Anchorage Jail Complex.

On April 6, 2015, the following events occurred:

- Law enforcement removed instrument 100351 from service due to gas flow errors
- Law enforcement removed the breath hose from instrument 100374 as a result of a subject spitting in the breath hose.
- Law enforcement replaced the removed breath hose with the breath hose from instrument 100351.

Instrument 100374, with the breath hose from 100351 installed, was in service until July 30, 2015.

The breath alcohol program staff has concluded a review of all breath tests administered on instrument 100374, with the breath hose from 100351 installed, and determined that all of the tests are scientifically valid and accurate.

A handwritten signature in black ink, appearing to read "Charles Foster".

Charles Foster
Scientific Director
Breath and Blood Alcohol Testing Program