

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100682

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 07/24/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.02: 0.080  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 1163 psi

BLANK TEST                      0.000    13:44  
INTERNAL STANDARD              ERROR    13:44

Internal Standard Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C              PASSED  
Breath Tube      = 48.1°C              PASSED

### PUMP INFO

Flow Rate = 4.765 L/M                  PASSED

### DETECTOR INFO

PUMP ON                                  PASSED  
PUMP OFF                                 PASSED

### FILTER INFO

Filter 1                                   PASSED  
Filter 2                                   PASSED  
Filter 3                                   PASSED

### INTERNAL STANDARD

PASSED