

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 9/16/2016 Instrument Serial Number: 100669

Instrument Location: Booking Room

Supervisor Name and Agency: Michael Fure / Yakutat Police

Describe the Status Change or Issue with the Instrument:

Time card is possibly bad. The time on it keeps changing about 2 hrs. off.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 9/16/2016 ~~14:22~~ 14:22

For Use by SCDL

Additional Notes

Email completed form to _____

For questions contact the Breath Alcohol Program at 907-269-5740