

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100398

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/29/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.87: 0.080
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 1146 psi

| | | |
|-------------------|--------------|-------|
| BLANK TEST | 0.000 | 02:37 |
| INTERNAL STANDARD | VERIFIED | 02:37 |
| EXTERNAL STANDARD | INTERFERENCE | 02:37 |

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 48.7°C | PASSED |
| Breath Tube = 45.2°C | PASSED |

PUMP INFO

| | |
|-----------------------|--------|
| Flow Rate = 4.441 L/M | PASSED |
|-----------------------|--------|

DETECTOR INFO

| | |
|----------|--------|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

INTERNAL STANDARD

PASSED