

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100378

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 02/01/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.38: 0.079  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 684 psi

BLANK TEST                    0.000   00:30  
INTERNAL STANDARD        VERIFIED   00:31

Gas Flow Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.1°C        PASSED  
Breath Tube       = 48.1°C       PASSED

### PUMP INFO

Flow Rate = 4.619 L/M        PASSED

### DETECTOR INFO

PUMP ON                        PASSED  
PUMP OFF                       PASSED

### FILTER INFO

Filter 1                        PASSED  
Filter 2                        PASSED  
Filter 3                        PASSED

INTERNAL STANDARD            PASSED