

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 04/03/2018
Effective: 04/04/2018

Version: DMEF2018 R0
Status: Active

Instrument # 100679

Date: 02/26/20

Analyst: Dereen Walton

1. Reason for instrument evaluation

Preventive maintenance due

2. Visual Inspection

Clean

Replace missing screws

Cermetex Box is secure

Ensure modem cable present

Gas Valve upright or secured to wall

All tubing and connections are secure

3. O-ring inspection and ensure 2 O-rings in replacement bag

4. Troubleshooting/Repair

Replaced o-ring. Routed filter wires separately.

Ran VOC and breath test normally.

Replaced on/off switch.

5. VOC (attached)

6. Internal Standard value $\leq 3.00\%$ Yes No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100679

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/26/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.26: 0.078
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1173 psi

BLANK TEST	0.000	10:46
INTERNAL STANDARD	✓ VERIFIED	10:47
Xq = 0.100 (0.18%)		
EXTERNAL STANDARD	0.078	10:47
X[1] = 0.0781 (-0.0015) (0.0001)		
BLANK TEST	0.000	10:48
EXTERNAL STANDARD	0.079	10:48
X[1] = 0.0788 (-0.0015) (0.0003)		
BLANK TEST	0.000	10:49
EXTERNAL STANDARD	0.079	10:49
X[1] = 0.0789 (-0.0015) (0.0002)		
BLANK TEST	0.000	10:50
EXTERNAL STANDARD	0.079	10:50
X[1] = 0.0787 (-0.0017) (-0.0004)		
BLANK TEST	0.000	10:51
EXTERNAL STANDARD	0.079	10:51
X[1] = 0.0786 (-0.0019) (-0.0001)		
BLANK TEST	0.000	10:52

Average = 0.0788
Std Dev = 0.0004

Diagnostic Check

VERSIONS ✓
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 48.1°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.83 V	PASSED
(1.00 V - 2.34 V)	
Cooler Voltage = 1.73 V	PASSED
(1.00 V - 2.32 V)	
Bias Voltage = 79 V	PASSED
(47 V - 111 V)	
Chopper Freq = 520 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.3 in	

PUMP INFO

Flow Rate = 4.086 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.001 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD ✓
Xq = 0.099 0.35% (0.00% - 4.00%)

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100679

Date: 02/26/2020
Time: 11:03:07

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
NON-DRINKING SUBJECT
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.27: 0.078
LOT #: AG826101
EXPIRATION: 09/18/2020

DIAGNOSTIC CHECK	PASSED	11:04
BLANK TEST	0.000	11:04
INTERNAL STANDARD	VERIFIED	11:04
EXTERNAL STANDARD	0.079	11:04
BLANK TEST	0.000	11:05
SUBJECT SAMPLE (Vol=3.35L)	0.000	11:05
BLANK TEST	0.000	11:06
EXTERNAL STANDARD	0.079	11:07
BLANK TEST	0.000	11:07

DW

