

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100697

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/27/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.35: 0.078
LOT #: 09214080A2
EXPIRATION: 05/01/2016
TANK PRESSURE: 891 psi

BLANK TEST	0.000	08:27
INTERNAL STANDARD	VERIFIED	08:27
EXTERNAL STANDARD	0.077	08:27
BLANK TEST	0.000	08:28
EXTERNAL STANDARD	0.077	08:28
BLANK TEST	0.000	08:29
EXTERNAL STANDARD	0.077	08:29
BLANK TEST	0.000	08:30
EXTERNAL STANDARD	0.077	08:30
BLANK TEST	0.000	08:31
EXTERNAL STANDARD	0.077	08:32
BLANK TEST	0.000	08:32

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED
Breath Tube = 46.7°C PASSED

PUMP INFO

Flow Rate = 4.389 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED


INTERNAL STANDARD

PASSED

COB 4/18/16


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

5/12/16

Subscribed and sworn before me this 12th day of May, 2016



Nikki Roth, Notary Public
My Commission Expires With Office

