

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100392

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/25/2014

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.91: 0.079  
LOT #: 11313080A1  
EXPIRATION: 06/01/2015

BLANK TEST	0.000	08:37
INTERNAL STANDARD	VERIFIED	08:37
EXTERNAL STANDARD	0.079	08:37
BLANK TEST	0.000	08:38
EXTERNAL STANDARD	0.079	08:38
BLANK TEST	0.000	08:39
EXTERNAL STANDARD	0.079	08:39
BLANK TEST	0.000	08:40
EXTERNAL STANDARD	0.079	08:40
BLANK TEST	0.000	08:41
EXTERNAL STANDARD	0.079	08:41
BLANK TEST	0.000	08:42

Average = 0.0790  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 2.00  
PIC: 3.02  
Modem: 2.3  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 48.0°C PASSED

### PUMP INFO

Flow Rate = 4.458 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

12/2/2014

Subscribed and sworn before me this 2nd day of Dec, 20 14

  
Nikki Roth, Notary Public  
My Commission Expires With Office



CPF  
12/1/14

