

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100688

Received From Hooper Bay

Date 6/29/17

Analyst Brandi Barnett

1. Reason for instrument return:

Detector overflow, (keyboard not working +  
modem not working at check in).

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload  Won't upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs  unable to check since won't upload

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)