

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 10/24/2025

Instrument #: 100345

Instrument Location: St Marys

Supervisor Name and Agency: Jared Michels AST

Describe the Status Change or Issue with the Instrument:

Instrument was having issues starting up and saying unable to load ticket base. Instrument was replaced with Serial #100379

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 10/24/2025 1200

Out of Service Date/Time: 10/24/2025 1200

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

100345-2025-10-24-Change-In-Status