

VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument
State of Alaska*

Serial #: 100375

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:03/11/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 27.18: 0.073
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 466 psi

BLANK TEST	0.000	15:33
INTERNAL STANDARD	VERIFIED	15:33
EXTERNAL STANDARD	INTERFERENCE	15:33

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber =	48.6°C	PASSED
Breath Tube =	42.4°C	PASSED

PUMP INFO

Flow Rate =	4.271 L/M	PASSED
-------------	-----------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED