

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 3/1/2018 Instrument Serial Number: 100410

Instrument Location: St. Mary's

Supervisor Name and Agency: Trooper Massey, AST

Describe the Status Change or Issue with the Instrument:

The instrument was unplugged during the January 2018 scheduled VOC period.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: ~~3/2018~~ 2/22/18 9:11 am BMB 3/7/18

Out of Service Date/Time: 1/2018

For Use by SCDL

Additional Notes

Instrument was unable to be uploaded for a period of time because it was turned off. The 1/3/18 voc was scheduled while the instrument was off. The 12/12/17 voc occurred as expected so the instrument was on then. There were no tests run between 12/12/17 and when the instrument was turned back on and put in service with a VOC on 2/22/18.

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740