

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 9/19/17 Instrument Serial Number: 100346

Instrument Location: Fairbanks Airport PD

Supervisor Name and Agency: Colleen O'Bryant, SCDL

Describe the Status Change or Issue with the Instrument:

I stopped by Fairbanks Airport PD to troubleshoot the modem line with Vance Inderte. The line was functional and abruptly stopped working. I opened the DMT cover and observed a slightly loose connection from the controller board to the modem board via the red/black two wire connector. I tightened the connection and checked all other connections, and restarted the instrument. DMT uploaded @ SCDL 9/18/17. CRB 9/19/17

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740