

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/25/21 Instrument Serial Number: 100405

Instrument Location: Crime Lab (Bear)

Supervisor Name and Agency: Derek Walton - SCDL

Describe the Status Change or Issue with the Instrument:

Checked instrument for VOC day and found no printed VOC.
Observed tank pressure to be 29 psi. Removed old tank, replaced
O-ring, and installed new tank (665 psi). VOC completed normally.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

| For Use by SCDL | Additional Notes |
|-----------------|------------------|
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740