

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 8/5/2025

Instrument #: 100361

Instrument Location: AST/Delta

Supervisor Name and Agency: C. O'Bryant

Describe the Status Change or Issue with the Instrument:

Sgt. Stuart contacted the breath section to describe the issues he encountered in an attempt to put DMT 100361 into service. The instrument was having issues with the screen connectivity as well as internal standard errors. Sgt. Stuart attempted to troubleshoot the screen with me on the phone. He turned the instrument off, disconnected power, disconnected and reconnected the screen from the SI board and returned power to the instrument. This did not alleviate the screen peeping and lack of visual function.

b COB 2025.08.26

100361 was never successfully placed in service at AST/Delta during this time frame. A successful In Service VOC was never accomplished.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Instrument being replaced due to screen and reported internal standard errors.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

100361-2025-08-05-Change-In-Status Page 1 of 1 COB 2025.08.26

~~100361-2025.08.06 Change In Status Page 1 of 1~~ COB 2025.08.26