

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100666

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/14/2014

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.44: 0.078
LOT #: 09214080A2
EXPIRATION: 05/01/2016

BLANK TEST	0.000	03:21
INTERNAL STANDARD	VERIFIED	03:21
EXTERNAL STANDARD	0.077	03:21
BLANK TEST	0.000	03:22
EXTERNAL STANDARD	0.077	03:22
BLANK TEST	0.000	03:23
EXTERNAL STANDARD	0.076	03:24
BLANK TEST	0.000	03:24
EXTERNAL STANDARD	0.076	03:25
BLANK TEST	0.000	03:26
EXTERNAL STANDARD	0.077	03:26
BLANK TEST	0.000	03:27

Average = 0.0766
Std Dev = 0.0005

Diagnostic Check

VERSIONS
DMT: 2.00
PIC: 3.03
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 42.8°C PASSED

PUMP INFO

Flow Rate = 4.828 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:



- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Nita J. Bolz
Scientific Director
State Breath Alcohol Program

12/18/2014

Subscribed and sworn before me this 18th day of Dec. 20 14


Nikki Roth, Notary Public
My Commission Expires With Office




12/18/14