

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Serial #: 100693

Date: 07/21/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.98: 0.080
LOT #: 01315080A3
EXPIRATION: 02/05/2017
TANK PRESSURE: 714 psi

BLANK TEST	0.000	10:31
INTERNAL STANDARD	VERIFIED	10:31
EXTERNAL STANDARD	0.076	10:31
BLANK TEST	0.000	10:32
EXTERNAL STANDARD	0.076	10:32
BLANK TEST	0.000	10:33
EXTERNAL STANDARD	0.076	10:33
BLANK TEST	0.000	10:34
EXTERNAL STANDARD	0.076	10:34
BLANK TEST	0.000	10:35
EXTERNAL STANDARD	0.075	10:35
BLANK TEST	0.000	10:36

Average = 0.0758
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 41.0°C PASSED

PUMP INFO

Flow Rate = 4.662 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

File Only: Instrument removed
from service. BMB 8/2/16