

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 6-3-20 Instrument Serial Number: 100674

Instrument Location: TOK

Supervisor Name and Agency: Christopher Bitz TOK AST

Describe the Status Change or Issue with the Instrument:

LCD Screen Does not function properly.
It turns on but images on screen are distorted
and not readable

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 6-3-20 @ 1530 hrs

For Use by SCDL

Additional Notes

Instrument was not placed into service at agency due to screen issue. DJW 06/03/2020

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740