

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100691

Received From Intoximeters

Date 08/17/21

Analyst Derek Watton

1. Reason for instrument return:

Warranty repair for screen freezing on startup

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

*NA - Instrument was not in the field
JMR 08/17/21*

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)

SRO Number: SRO-046018

Received Date: 05/17/2021

Completion Date: 08/02/2021

SRO Type: REPAIR

SRO Description: DMT Repair/Preventive Maintenance

Customer ID: C000003651

Customer PO: WTY03

Contact: Derek Walton

Email: derek.walton@alaska.gov

Ship To:

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507

Units on SRO

1	LEGDMT	LEGACY DMT
---	--------	------------

Service Line: 1		**All Instruments are Calibrated to Factory Specifications**	
Unit Type:	(LEGDMT) LEGACY DMT		
Serial Number:	50100691		
Parts:	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	2	59-0058-00	DMT ELEC IC RS232 MULTCH U31/33
Reason:	DMT Functional		
	Display		
Reason Notes:	Customer Reported Issue: Screen freezing		
Resolution:	Replaced		
	RPL PCB		
Resolution Notes:	Replaced bad Green SI Board. Installed Alaska 43402 software. Performed Calibration/Adjustment, Wet and Dry Gas Accuracy Checks, Breath Tests, and outgoing Diagnostics Test. Verified Modem with laptop and DMT Tech Software.		
	1	59-0076-00	DMT ELEC PCB SI KIT 43602

Order Packing Slip Report

Intoximeters
Experience · Service · Integrity

8/3/2021 12:41:10 PM

Packing Slip: 233754

Pack Date: 8/3/2021

SRO: SRO-046018



From: Intoximeters, Inc
2081 Craig Road
St Louis MO 63146
USA

Customer: C000003651

Ship To: 2

Alaska Department of Public Safety
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Alaska Department of Public Safety
Procurement & Supply Section
4805 Dr. MLK Jr. Avenue
Anchorage AK 99507
USA

Customer PO: WTY03

Ship Via: FDX

FED EX

Weight: 37.00

Delivery Terms: 2D

Shipping Terms: DESTINATION

Packages: 1

Customer Ship Account:

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100691
1	59-0058-00	EA		2.00	0.00
	DMT ELEC IC RS232 MULTCH U31/33				
2	59-0076-00	EA		1.00	0.00
	DMT ELEC PCB SI KIT 43602				