

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100413

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/14/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.12: 0.078
LOT #: 01315080A2
EXPIRATION: 02/05/2017
TANK PRESSURE: 1015 psi

BLANK TEST	0.000	16:06
INTERNAL STANDARD	VERIFIED	16:06
EXTERNAL STANDARD	0.076	16:06
BLANK TEST	0.000	16:07
EXTERNAL STANDARD	0.075	16:07
BLANK TEST	0.000	16:08
EXTERNAL STANDARD	0.076	16:08
BLANK TEST	0.000	16:09
EXTERNAL STANDARD	0.076	16:10
BLANK TEST	0.000	16:11
EXTERNAL STANDARD	0.075	16:11
BLANK TEST	0.000	16:12

Average = 0.0756
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 47.8°C PASSED

PUMP INFO

Flow Rate = 5.045 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



Charles R. Foster
Scientific Director
State Breath Alcohol Program

4/11/17

Subscribed and sworn before me this 11th day of April, 20 17



Nikki Roth, Notary Public
My Commission Expires With Office




SMB
4/11/17