

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 10/16/2023

Instrument #: ~~100687~~ 100697

COB 2023.11.16

Instrument Location: Anchorage Police Department

Supervisor Name and Agency: Colleen O'Bryant - SCDL

Describe the Status Change or Issue with the Instrument:

I reached out to BTS Nicholas Bowe regarding diagnostic failed status messages due to breath tube temperature being too low. BTS Bowe inspected the breath hoses and reconnected to ensure a secure fitting of the heating element. The breath tube temperature increased. The instrument was never taken out of service and remains in service as of 2023.10.16.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

See attached email conversations. COB 2023.10.16

*Updated to add header and footer information. COB 2024.01.16

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

~~100697 2023.10.16 Change In Status~~ ^{COB} Page 1 of 5

100697 2023.10.16 Change In Status - Amended.

COB 2/13/24

Date: 10/16/2023

Instrument #: ~~100687~~ 100697
COB 2023.11.16

Instrument Location: Anchorage Police Department

Supervisor Name and Agency: Colleen O'Bryant - SCDL

Describe the Status Change or Issue with the Instrument:

I reached out to BTS Nicholas Bowe regarding diagnostic failed status messages due to breath tube temperature being too low. BTS Bowe inspected the breath hoses and reconnected to ensure a secure fitting of the heating element. The breath tube temperature increased. The instrument was never taken out of service and remains in service as of 2023.10.16.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

See attached email conversation. COB 2023.10.16

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

100697 2023.10.16 Change in Status - Amended

~~100697 2023.10.16 Change In Status~~ Page 2 of 5

100697 2023.10.16 Change In Status Page 1 of 4

COB 2/13/24

Date: 10/16/2023

Instrument #: 100687

Instrument Location: Anchorage Police Department

Supervisor Name and Agency: Colleen O'Bryant - SCDL

Describe the Status Change or Issue with the Instrument:

I reached out to BTS Nicholas Bowe regarding diagnostic failed status messages due to breath tube temperature being too low. BTS Bowe inspected the breath hoses and reconnected to ensure a secure fitting of the heating element. The breath tube temperature increased. The instrument was never taken out of service and remains in service as of 2023.10.16.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

See attached email conversation. COB 2023.10.16

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

100097 2023.10.16 change in status - Amended

~~100697 2023.10.16 Change In Status~~ ^{COB} Page 3 of 5

100697 2023.10.16 Change In Status Page 2 of 4

COB 10/13/24

O'Bryant, Colleen S (DPS)

From: Bowe, Nicholas <nicholas.bowe@anchorageak.gov>
Sent: Thursday, October 12, 2023 10:17 PM
To: O'Bryant, Colleen S (DPS)
Subject: RE: Dry Gas Request + 100687, 100697 Breath Tube Temperature

I had tightened the plug on the breath hose for instrument 100697 yesterday and today it looks like it is in the operating range for the temp (39+). I tightened the plug on the hose for 100687 today, and the temp started climbing and is now in operational range (39+) as well. I will not take them out of service at this time. – Thanks – N. Bowe

From: O'Bryant, Colleen S (DPS) <colleen.obryant@alaska.gov>
Sent: Thursday, October 12, 2023 7:01 AM
To: Bowe, Nicholas <nicholas.bowe@anchorageak.gov>
Cc: Walton, Derek J (DPS) <derek.walton@alaska.gov>; Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>
Subject: Dry Gas Request + 100687, 100697 Breath Tube Temperature

[EXTERNAL EMAIL]

Nick,

We can definitely get you new tanks. While reviewing your instruments I noticed that 100687 failed diagnostic tests 4 times in the last week due to the breath tube temperature being too low. 100697 also failed once on 2023.10.07 for the same reason.

Could you please visit these instruments and unplug the breath tube heating connection, reconnect it and make sure it is seated properly. The connection should be fully inserted into the back of the instrument and then snugly screwed in. It's also important to make sure the temperature in the room the instrument lives isn't getting too cold. If this doesn't bring up your breath tube temperature we'll need to replace these instruments.

Colleen O'Bryant

Forensic Scientist III - Chemistry
Alaska Scientific Crime Detection Laboratory
4805 Dr. Martin Luther King Jr. Ave
Anchorage, Alaska 99507

907-269-5783 (direct)
907-350-2111 (cell)

From: Bowe, Nicholas <nicholas.bowe@anchorageak.gov>
Sent: Wednesday, October 11, 2023 8:04 PM
To: DPS SCDL Toxicology (DPS sponsored) <dps.scdl.toxicology@alaska.gov>
Subject: dry gas

CAUTION: This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

100697 2023.10.16 Change in status - Amended
~~100697 2023.10.16 Change In Status - Page 4 of 5~~

CAB 2/13/24

Hi there! I was wondering if I could swing by some time (probably next week) and pick up some new dry gas bottles for the Datamasters at the Anchorage jail. I would probably be wanting 3 - just to make sure we make it through the holidays + an upcoming academy training. We are down to 1 as of last night, but I will probably change out one of them tonight because one of the Datamasters is in the red. Thanks - N. Bowe

100697 2023.10.16 Change in Status - Amended.

COB

~~100697 2023.10.16 Change In Status Page 5 of 5~~

100697 2023.10.16 Change In Status Page 4 of 4

COB 9/13/24