



State of Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form - DMCAF2011R1

Issuing Authority - Nita Bolz, Scientific Director

Approval date: 03/10/11

Effective date: 03/10/11

Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100317

Date: 3/22/11

*Supervised cadework  
with C. Bryant  
COB 3/22/11*

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Guth
- c. Lot 9380
- d. Expiration Date 12/17/11

2. Wet Bath Simulator

- a. Serial Number DR4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100347  
-----

Date: 03/22/2011  
Time: 14:26:48

SUPERVISOR NAME:  
CHARLES FOSTER  
CERTIFICATE #: 7734

LOT #: GUTH100  
BOTTLE #: 9380  
EXPIRATION: 12/17/2011

Ca = 0.1000  
CAL = 0.967916    0.800 <= CAL < 1.200  
b1 = 0.0002    0.0000 <= b1 < 0.0040  
b2 = 0.0053    0.0010 <= b2 < 0.0100  
b3 = 0.0000    0.0000 <= b3 < 0.0040  
Xq = 0.0921    0.0500 <= Xq < 0.2000  
a21 = 1.175203    1.050 <= a21 < 1.300  
a31 = 0.445516    0.300 <= a31 < 0.800