

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/06/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.97: 0.077
LOT #: AG335202
EXPIRATION: 12/18/2025
TANK PRESSURE: 585 psi

BLANK TEST	0.000	22:47
INTERNAL STANDARD	VERIFIED	22:47
EXTERNAL STANDARD	0.076	22:47
BLANK TEST	0.000	22:48
EXTERNAL STANDARD	0.076	22:48
BLANK TEST	0.000	22:49
EXTERNAL STANDARD	0.076	22:50
BLANK TEST	0.000	22:50
EXTERNAL STANDARD	0.076	22:51
BLANK TEST	0.000	22:52
EXTERNAL STANDARD	0.076	22:52
BLANK TEST	0.000	22:53

Average = 0.0760
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C	PASSED
Breath Tube = 44.5°C	PASSED

PUMP INFO

Flow Rate = 4.514 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only DM 06/26/25