

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100346

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:03/31/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.15: 0.081
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 1082 psi

BLANK TEST	0.000	16:04
INTERNAL STANDARD	VERIFIED	16:04
EXTERNAL STANDARD	0.090	16:04

Standard Out of Range

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 47.9°C	PASSED

PUMP INFO

Flow Rate = 4.650 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

INTERNAL STANDARD	PASSED
-------------------	--------