

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100378

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/04/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.30: 0.078
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 1242 psi

BLANK TEST	0.000	19:59
INTERNAL STANDARD	VERIFIED	19:59
EXTERNAL STANDARD	0.078	19:59
BLANK TEST	0.000	20:00
EXTERNAL STANDARD	0.077	20:00
BLANK TEST	0.000	20:01
EXTERNAL STANDARD	0.077	20:01
BLANK TEST	0.000	20:02
EXTERNAL STANDARD	0.076	20:02
BLANK TEST	0.000	20:03
EXTERNAL STANDARD	0.077	20:03
BLANK TEST	0.000	20:04

Average = 0.0770
Std Dev = 0.0007

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED
Breath Tube = 48.4°C PASSED

PUMP INFO

Flow Rate = 4.493 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO


Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

8/10/18

Subscribed and sworn before me this 10 day of 08, 2018


Charlie K. Bailey, Notary Public
My Commission Expires With Office



JM 07/19/18