

Alaska Scientific Crime Detection Laboratory

DataMaster Acceptance Form

Issued: 01/27/2012
Effective: 01/27/2012

Version: DMAF2012 R0
Status: Active


Instrument # 100687

Date: 5/11/12

Analyst: CTCF

- Check in instrument in to JusticeTrax Breath Alcohol Database
- Visual Inspection
- Update software *USB Already Fixed*
- Print calibration factors (attached)
- NPAS calibration paperwork (attached)
- Barcode Instrument
- State of Alaska Property ID# 12 - F0378

Already vented to seat, controller board new



CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100687

Date: 10/24/2011
Time: 09:24:54

Ca = 0.1000
CAL = 1.019002 0.800 <= CAL < 1.200
b1 = 0.0009 0.0000 <= b1 < 0.0040
b2 = 0.0040 0.0010 <= b2 < 0.0100
b3 = 0.0021 0.0000 <= b3 < 0.0040
Xq = 0.1354 0.0500 <= Xq < 0.2000
a21 = 1.149434 1.050 <= a21 < 1.300
a31 = 0.503196 0.300 <= a31 < 0.800



Certificate of Calibration

100687

Instrument Serial Number

10260

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

10/24/11

Date

A handwritten signature in black ink, appearing to read "C. Foster", is written over a horizontal line.

Technician

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100687\B-100687 ECF-20140

7/10/2014