

In Field Instrument Review for 5/30/2016 VOC Day

Dates 4/11/2016-5/30/2016

Prepared By: Brandi Barnett

100345- AST/Soldotna – No Action Taken

- RFI (1), Blank Error (1), Standard Out of Range (1), Ambient Fail (3)- 24 tests total

100346- Fairbanks Airport PD – No Action Taken

- Interference (1), Blank Error (1) – 10 tests total

100347- Cordova PD – No Action Taken

- Gas Flow Error (1) – 11 tests total

100354- BBBPD – VOC reading 0.005 low. Appears to be tank concentration decreasing. Replace tank

100355- Seward PD – No Action Taken

- RFI (2) – 19 tests total

100357- St Paul – Instrument was out of service from 5/1/2016 until 6/8/2016 due to tank expiration.

100364- Anchorage Jail – No Action Taken

- Blank Error (3) – 97 tests total

100366- Kenai PD – Removed from Service on 6/9/2016

- Pump Error (1), Standard Out of Range (5) – 15 tests total
- Diagnostic Failed on 5/30/2016 VOC

100370- Homer PD – No Action Taken

- Interference (1) – 7 tests total

100371- Glacier Bay – VOC reading 0.005 low. Appears to be tank concentration decreasing. Replace tank.

100373- AST Nenana – No Action Taken

- Gas Flow Error (1) – 3 tests total

100375- Seldovia – Instrument went out of service on 5/1/2016 due to tank expiration, no current operators so instrument being removed from service.

100395- Anchorage Jail – No Action Taken

- RFI (5), Pump Error (2), Interference (1), Standard Out of Range (1) – 103 tests total

100400- Palmer PD – No Action Taken

- RFI (6), Pump Error (1), Gas Flow Error (2), Ambient Fail (1) – 67 tests total
- Spoke with Luke Szipsky about RFI because it was mentioned by another officer in the valley and determined it was likely an operator issue.

100404- Anchorage Jail – No Action Taken

- RFI (1), Pump Error (2), Interference (1), Blank Error (1) – 54 tests total

100414- AST Emmonak –Removed from Service on 4/7/2016 for preventative maintenance

100417- Juneau PD – Replacement instrument being sent during week of 6/20

- Suck Back Error (2) – 18 tests total (had one suck back error last time)

100420- Ft Greely – Instrument was out of service from 5/1- 6/10 due to tank expiration

100422- JBER – No Action Taken

- Pump Error (1) – 19 tests total

100423- Ft Wainwright – No Action Taken

- RFI (1), Pump Error (1) – 16 tests total

100665- AST St. Mary – No Action Taken

- Standard Out of Range (3) – 6 tests total (All were prior to tank being switched on 4/27/2016)

100667- AST Matsu West – No Action Taken

- RFI (3) – 17 tests total

100671- AST Fairbanks – No Action Taken

- RFI (1) – 26 tests total

100673- NSBPD – No Action Taken

- Pump Error (1), Standard Out of Range (1)- 4 tests total (SOR was prior to tank change on 5/4/2016)

100677- Bethel PD- Removed from service on 4/21/2016 for modem connectivity problems

100678- AST Anchor Point – No Action Taken

- Pump Error (1), Interference (1), Tank Interference (1)- 16 tests total

100682- Kake PD – Out of Service, waiting for return

100689- Kotzebue PD (Upload ended 5/19) – No Action Taken

- Tank Interference (1), Standard Out of Range (1)- 16 tests total (SOR was prior to tank change on 5/3)

100690- Sitka PD – No Action Taken

- Interference (1) – 10 tests total

100700- Soldotna PD – No Action Taken

- RFI (2), Tank Interference (1) – 19 tests total



