

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100376

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/26/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 28.83: 0.077

LOT #: ~~AG-018102~~ AG-018102 CUG-1118/21 *A*

EXPIRATION: 06/29/2022

TANK PRESSURE: 1174 psi

BLANK TEST	0.000	11:40
INTERNAL STANDARD	VERIFIED	11:40
EXTERNAL STANDARD	0.076	11:41
BLANK TEST	0.000	11:41
EXTERNAL STANDARD	0.076	11:42
BLANK TEST	0.000	11:42
EXTERNAL STANDARD	0.076	11:43

User Abort

** Incorrect lot #.
Sec 100376 2021.11.08 change in status
CUG-1119/21*

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED

Breath Tube = 46.3°C PASSED

PUMP INFO

Flow Rate = 3.840 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED