

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100410

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/30/2025

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.61: 0.079  
LOT #: AG335202  
EXPIRATION: 12/18/2025  
TANK PRESSURE: 110 psi

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | 0.000    | 17:01 |
| INTERNAL STANDARD | VERIFIED | 17:01 |
| EXTERNAL STANDARD | 0.078    | 17:01 |
| BLANK TEST        | 0.000    | 17:02 |
| EXTERNAL STANDARD | 0.078    | 17:02 |
| BLANK TEST        | 0.000    | 17:03 |
| EXTERNAL STANDARD | 0.078    | 17:04 |
| BLANK TEST        | 0.000    | 17:04 |

Tank Pressure Too Low

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

|                         |        |
|-------------------------|--------|
| Sample Chamber = 48.6°C | PASSED |
| Breath Tube = 47.2°C    | PASSED |

### PUMP INFO

|                       |        |
|-----------------------|--------|
| Flow Rate = 4.587 L/M | PASSED |
|-----------------------|--------|

### DETECTOR INFO

|          |        |
|----------|--------|
| PUMP ON  | PASSED |
| PUMP OFF | PASSED |

### FILTER INFO

|          |        |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

|                   |        |
|-------------------|--------|
| INTERNAL STANDARD | PASSED |
|-------------------|--------|

File Only JMW 06/18/25