

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 6/19/19 Instrument Serial Number: 100384

Instrument Location: Hoonah

Supervisor Name and Agency: Sgt. Joshua Huskey Hoonah PD

Describe the Status Change or Issue with the Instrument:

Instrument gave double beeps when pressing any button.
Also reads 0 psi on a relatively new bottle. While turned
on instrument is very warm to the touch.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 6/5/2019 1800 hrs.

For Use by SCDL

Additional Notes

Sample chamber temp 0.0°C BMB 9/6/19

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740