

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100375

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 03/11/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 27.18: 0.073  
LOT #: AG608504  
EXPIRATION: 03/25/2018  
TANK PRESSURE: 462 psi

BLANK TEST	0.000	15:38
INTERNAL STANDARD	VERIFIED	15:39
EXTERNAL STANDARD	INTERFERENCE	15:39

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.5°C	PASSED
Breath Tube = 43.1°C	PASSED

### PUMP INFO

Flow Rate = 4.417 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED