

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100375

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/07/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 27.80: 0.074
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 1138 psi

BLANK TEST	0.000	16:19
INTERNAL STANDARD	VERIFIED	16:19
EXTERNAL STANDARD	0.075	16:19
BLANK TEST	0.000	16:20
EXTERNAL STANDARD	0.072	16:20
BLANK TEST	0.000	16:21
EXTERNAL STANDARD	0.071	16:21

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C	PASSED
Breath Tube = 44.9°C	PASSED

PUMP INFO

Flow Rate = 4.451 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------