

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100406

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/26/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 27.38: 0.073
 LOT #: AG135004
 EXPIRATION: 12/16/2023
 TANK PRESSURE: 787 psi

BLANK TEST	0.000	09:31
INTERNAL STANDARD	VERIFIED	09:31
EXTERNAL STANDARD	0.073	09:31
BLANK TEST	0.000	09:32
EXTERNAL STANDARD	0.073	09:32
BLANK TEST	0.000	09:33
EXTERNAL STANDARD	0.072	09:33
BLANK TEST	0.000	09:34
EXTERNAL STANDARD	0.072	09:34
BLANK TEST	0.000	09:35
EXTERNAL STANDARD	0.072	09:35
BLANK TEST	0.000	09:36

Average = 0.0724
 Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
 PIC: 3.02
 Modem: 2.6
 Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 40.7°C	PASSED

PUMP INFO

Flow Rate = 4.541 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED


FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

I, Derek J. Walton, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) As of the date signed below, I certify that the calibration of the referenced instrument was accurate on the date in which the above tests were performed and therefore certified for evidentiary use in the State of Alaska.



 Derek J. Walton
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 21 day of 11, 2023 ^{11 PC 11/21/23}



 Notary Public
 My Commission Expires With Office



Handwritten: 09/26/23