

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 1/4/2013
Effective: 1/4/2013

Version: CIISF2013 R0
Status: Active

Date: June 2, 2015 Instrument Serial Number: 100691

Instrument Location: Glacier Bay NP&P

Supervisor Name and Agency: Wendy Bredow - Glacier Bay NP&P - National Park Service

Describe the Status Change or Issue with the Instrument:

The DMT was placed in "Out of Service" status. After a tank change, this instrument failed a VOC due to a "Detector Overflow". A second VOC was attempted resulting was the same "Detector Overflow" message. Additionally, the screen saver has seized up and is no longer moving about the screen.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: June 2, 2015, Approx. 1930 hrs.

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740