



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 1232
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 02511080A1
Expiration: 2/1/2013

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAWASTER cdm 130156

MARCH 01, 2011

OPERATOR'S NAME:
DLN
OPERATOR'S NUMBER: 7732
ALCO TARGET VALUE: 0.080
ALCO S/N: 88A1

0.080 BAC (For use with breath alcohol testing instruments)


--- SUPERVISOR MODE ---

Contents: 105 Liters @ 1025 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	208.4 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Balance	BAC whichever is greater	

.080 ADJUSTED FOR 30.06 in		
ALCO TARGET	.080	14:05
BLANK TEST	.000	14:06
INTERNAL STANDARD	VERIFIED	14:06
ALCO TV 30.06 in	.083	14:06
BLANK TEST	.000	14:07
ALCO TV 30.06 in	.081	14:08
BLANK TEST	.000	14:08
ALCO TV 30.06 in	.081	14:09
BLANK TEST	.000	14:09
ALCO TV 30.06 in	.082	14:10
BLANK TEST	.000	14:10
ALCO TV 30.06 in	.081	14:11
BLANK TEST	.000	14:11

N = 5
S.D. = .0003
AVG. = .0816
S.D. = .0008
S.E. = 1.5147


Specialty Gas Lab Manager

1-28-2011
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

